



NOA.040.22
March 30, 2022

NOTICE OF AWARD

JOSE ALBERTO R. CATRAL
President
MEDICAL TRENDS & Technologies, Inc.
200 J. Elizalde St., Phase V, BF Homes
Parañaque City

Dear Mr. Catral:

This is to inform you that as a result of the competitive public bidding conducted on February 02, 2022 relative to the Project, **re: Supply and Delivery of Medical Supplies** is hereby awarded to you as the Lowest Calculated and Responsive Bid (LCRB) at bid offer amounting to **Php 7,648,000.00.**

Please acknowledge agreement to the contents of this **"NOTICE OF AWARD"** by signing under the word **"CONFORME"** below and return the original to us within five (5) working days from receipt hereof.


Also, you are hereby required to provide within ten (10) days the performance security in the form and the amount stipulated in the Instruction to Bidders. Failure to provide the performance security shall constitute sufficient ground for cancellation of the award and forfeiture of the bid security.

Please be guided accordingly.

Very truly yours,


JOEL M. ABANILLA, MD.
Executive Director

CONFORME:

for: 
JOSE ALBERTO R. CATRAL
President

Michelle T. Cruz, RN
Regional Sales Manager

IMPORTANT:

FAILURE TO RETURN THIS DOCUMENT WITHIN FIVE (5) WORKING DAYS FROM RECEIPT SHALL RESULT IN THE CANCELLATION OF THE AWARD. PLEASE SEND BACK TO BIDS & AWARDS COMMITTEE OFFICE IMMEDIATELY.



PHILIPPINE HEART CENTER
East Avenue, Quezon City

Abstract of Sealed Bidding: Medical Supplies, for the period of March 2022 to December 2022 (ITB.008.22)

MEDICAL TRENDS AND TECHNOLOGIES, INC.

Item#	SPECIFICATION	SIZE	UNIT OF MEASURE	TOTAL QTY.	UNIT COST	BID PRICE	BIDDER
66	BLOOD GAS AND ELECTROLYTE CARTRIDGES		piece	16,000	591.03	478.00	Medical Trends & Technologies
	must provide machine;						
	TECHNICAL SPECIFICATIONS FOR CARTRIDGES						
	1. Test must provide results for:						
	a. Blood Gas (pH, pCO2, pO2, HCO3, TCO2, PE, O2Sat)						
	b. Electrolytes (Na, K, iCa)						
	c. Hematology						
	d. Glucose						
	e. Lactate						
	2. The distributor will replace any defective cartridges that is being delivered to the end-user						
	3. Minimal volume of blood sample required (not more than 1 cc of blood sample)						
	TECHNICAL SPECIFICATIONS FOR ABG LOANER UNITS						
	1. The machine must be capable of determining the following tests:						
	a. Blood Gas						
	b. Electrolytes and Hematology						
	c. Chemistry						
	2. Must have the least variance from the main laboratory determination both cartridges and analyzer upon opening a new lot number of cartridges						
	4. Must provide an external simulator device kit that will test the quality control of the analyzer/or maintenance free and capable of performing quality control/calibration within the system						
	5. Keypad control						
	6. Capable of linking with data management system thru USB or network/or capable through network system						
	7. Reading time for test must not exceed 3 minutes						
	8. Must be battery operated						
	9. The distributor will provide a battery replacement to ensure a continuous machine operation						
	10. Must have at least LCD screen						
	11. The weight must not exceed 1.5 lbs.						
	12. Must provide at least four (4) ABG machines						
	13. Must include printer for result printing						
	14. Maintenance and replacement of parts should be the sole responsibility of the distributor						
	15. Must perform a free regular preventive maintenance checkup						
	16. The distributor must provide certificate of training of the technician who will perform the maintenance checkup						
	17. Response time in case of machine malfunction must be at least 2 hours upon receipt of calls						
	18. A readily available functional back-up unit in case of prolonged preventive maintenance						
	19. Must have a certificate of quality management standard compliance						

TOTAL BID : P7,648,000.00

CERTIFIED CORRECT

JEANETTE Z. BURILLO, RL
Chairman, BAC Secretariat