



Republic of the Philippines
PHILIPPINE HEART CENTER
East Avenue, Quezon City 1100, Philippines
Telephone Nos. 8925-2401 to 50 www.phc.gov.ph

NOA.046..24
January 01, 2024

NOTICE OF AWARD

JOSE ALBERTO R. CATRAL
President/Chairman of the Board
MEDICAL TRENDS & TECHNOLOGIES, INC.
200 J. Elizalde St., Phase V. BF Homes
Paranaque City


Dear Mr. Cabral:

This is to inform you that as a result of the Competitive Public Bidding conducted last October 18, 2023, relative to the **Project, re: Medical Supplies**, is hereby awarded to you as the Lowest Calculated and Responsive Bid (LCRB) at total bid offer amounting to **Php 10,713,600.00**.

Please acknowledge agreement to the contents of this **"NOTICE OF AWARD"** by signing under the word **"CONFORME"** below and return the original to us within five (5) working days from receipt hereof.

Also, you are hereby required to provide within ten (10) days the performance security in the form and the amount stipulated in the Instruction to Bidders. Failure to provide the performance security shall constitute sufficient ground for cancellation of the award and forfeiture of the bid security.

Very truly yours,


JOEL M. ABANILLA, M.D.
Executive Director/ Head of the Procuring Entity

CONFORME:


JOSE ALBERTO R. CATRAL
President/Chairman of the Board

1/5/2024
Date Signed

IMPORTANT:
FAILURE TO RETURN THIS DOCUMENT WITHIN FIVE (5) WORKING DAYS FROM RECEIPT SHALL RESULT IN THE CANCELLATION OF THE AWARD. PLEASE SEND BACK TO BIDS & AWARDS COMMITTEE OFFICE ASAP.



PHILIPPINE HEART CENTER
East Avenue, Quezon City


Abstract of Sealed Bidding: Medical Supplies, for the period of January 2024 to December 2024 (ITB.085.23)

MEDICAL TRENDS & TECHNOLOGIES

Item#	SPECIFICATION	UNIT OF MEASURE	TOTAL QTY.	UNIT COST	BID PRICE	BIDDER
1. OPERATING ROOM UNIT / PERFUSION SECTION						
1.4 MISC. MEDICAL SUPPLIES STOREROOM						
77	BLOOD GAS ANALYSIS CARTRIDGE must provide machine; TECHNICAL SPECIFICATIONS FOR CARTRIDGES 1. Test must provide results for: a. Blood Gas (pH, pCO ₂ , pO ₂ , HCO ₃ , TCO ₂ , PE, O ₂ Sat b. Electrolytes (Na, K, iCa) c. Hematology d. Glucose e. Lactate 2. The distributor will replace any defective cartridges that is being delivered to the end user TECHNICAL SPECIFICATIONS FOR ABG LOANER UNITS 1. The machine must be capable of determining the following tests: a. Blood Gas b. Electrolytes and Hematology c. Chemistry d. Coagulation e. Cardiac Marker 2. Keypad control (use of finger only and not stylus) 3. Capable of linking with data management system thru USB or network 4. Reading time for test must not exceed 3 minutes 5. Must be battery operated 6. The distributor will provide a battery replacement to ensure a continuous machine operation 7. Must have at lest LCD screen 8. The weight must not exceed 1.5 lbs 9. Must provide at least four (4) ABG machines 10. Must include printer for result printing 11. Maintenance and replacement of parts should be the sole responsibility of the distributor 12. Must perform a free regular preventive maintenance checkup 13. The distributor must provide certificate of training of the technician who will perform the maintenance checkup 14. A readily available functional back-up unit in case of prolonged preventive maintenance 15. Must have a certificate of quality management standard compliance	PIECE	21,600	520.80	496.00	Medical Trends (Istat CG8+)

TOTAL BID : P10,713,600.00

CERTIFIED CORRECT :


JEANETTE Z. BURILLO, RL
Chairman, BAC Secretariat

Handwritten initials