

PHILIPPINE HEART CENTER

East Avenue, Quezon City

PURCHASE ORDER

No. 0928-2013

TO: ZENITH MEDICAL EQUIPMENT, INC.
 (Name of Supplier)
1123 RIZAL AVENUE, MANILA
 (Address)

February 12, 2013
 Date

Please deliver to PHILIPPINE HEART CENTER, East Avenue, Quezon City

AND CHARGE TO OUR ACCOUNT THE FOLLOWING:

| QUANTITY | UNIT | SPECIFICATION | UNIT COST | AMOUNT |
|----------|------|---|-------------|---------------|
| 1 | unit | <p><i>depression</i></p> <p>HYPO/HYPERTHERMIA MACHINE Portable Thermal Machine Technical Specifications: 1. Cooling and warming system 2. Dual reservoir 3. Control panel (one touch selection and large buttons) with LED digital displays of the temperature (both patient and water) in Centigrade and Fahrenheit Temp. Control (display range) - Patient Temperature Range 30deg C to 41 deg. - Water Temp. Range: 4 deg C to 42 deg C 4. Visual and audible alerts 5. Integrated safety precaution (Dual over-temperature monitoring systems) with automatic shut off. 6. Accompanying Thermal pads (adult and pedia size) and re-usable rectal probe per machine. 7. Flow rate 16gph (60lph) minimum with blanket 8. Refrigerant: Freon Free (non-CFC) 9. Multiple modes for both cooling and heating. 10. Power requirements: 120 volts/60Hz. Accessories: 1. Thermal Pad Heavy duty reusable blanket, all polymer, adult 1- Thermal Pad Heavy duty reusable blanket, all polymer, pedia 1- Reusable rectal probe</p> <p><i>2/20/13</i></p> | P786,500.00 | P3,146,000.00 |

Funds Available

ROSALIO T. OCLARES
 Chief, Accounting Division

| | | |
|---|----------------------------|----------------------------------|
| PURPOSE PSD - Direct Issue Replenishment of Stock | BASIS Sealed Bid - 4.13 | OBLIGATION NO. 13-02-108 (co) |
|---|----------------------------|----------------------------------|

THIS PURCHASE ORDER IS MADE SUBJECT TO THE CONDITIONS AND SPECIAL INSTRUCTIONS PRINTED AND/OR TYPEWRITTEN AT THE BACK HEREOF.

Approved By: *[Signature]*
 MANUEL T. CHUA CHIACO, JR., M.D.

Executive Director