



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 015-2023

SUBJECT : Revised/Amended Technical Specifications

**PROJECT/S : 5 units Stretcher (Negotiated Procurement-
2 Failed Biddings) - ITB No. 006.23**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents, thus, revisions/changes are introduced.


Description of Function	
Stretcher is used for transfer or transport of patient for procedure and diagnostic test.	
Operational Requirements	
System complete with all accessories as specified in the Technical Specifications below.	
	Technical Specifications
1	Hydraulic height adjustment
2	Four (4) wheel casters with central locking system or its equivalent.
3	Weight capacity: safe working load 300kg
4	Height adjustment:
	4.1. High position: maximum 100 cm
	4.2. Lowest position: maximum 54 cm
5	Width: 100 cm
6	Length: 175 cm (bumpers are included in the measurement)
7	Side rails: fixable and collapsible with double locking system for patient's safety or its equivalent
8	Power assist fowler back rest
9	Mattress:
	9.1. washable and waterproof cover to keep mattress from fluid penetration
	9.2. thickness: 2 – 3 inches
10	Safety Belt:
	10.1. push button buckle type (body and lower extremities)
	10.2. Strap Type: coated polyester vinyl, fluid resistant and can be cleaned and sanitized
11	Compartment for patient's belongings
12	Control brake

13	Integrated portable oxygen holder, oblique (stainless)
14	Adjustable IV pole *height, with 2 – 4 hooks, removable or collapsible
15	Urine bag holder
16	With 2 years warranty for parts and services
17	Preventive maintenance twice a year for three (3) years
18	ISO 13485
19	Provision of user's manual both hard and soft copy upon delivery
20	Product demonstration of the same unit during post qualification.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 10th day of March 2023, BAC Office, PHC.


MARIETTA A. VENASCO, RN, MAN, MBAH
Chairman, BAC for Pharmaceutical Supplies & Med Equipment