



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 017-2023**

**SUBJECT : Revisions in the Technical Specifications**

**PROJECT/S : 1 unit Nitric Oxide Therapy Machine  
(Negotiated Procurement -2 Failed Biddings)  
- ITB No. 017.23**

\*\*\*\*\*

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents, thus, revisions/changes are introduced.

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended "Should be read as"
Page 7	1.8. Printer – integrated printer should be available to print times result trending and alarm warning messages	<b>1.8. USB Port for retrieving alarm warning messages</b>
	1.12. Calibration should be simple	<b>1.12. To be deleted</b>
Page 8	3.2. The unit shall be capable of operating in ambient temperature of 10-14 deg C and relative humidity of less than 15-90%	<b>3.2. The unit shall be capable of operating in ambient temperature of 10-15 deg C and relative humidity of less than 15-90%</b>
	5.4. Comprehensive warranty (3 years) & 5 years CMC after completion of warranty period	<b>5.4. Rates of consumables &amp; accessories should be frozen for 6 years</b>  <b>Scope of Supply</b> <b>- Single Use Patient Circuit-Kit</b> <b>- Nitric Oxide-Tank</b> <b>- Calibration Kit</b>
	5.5. Rates of consumables & accessories should be freezed for 8 years. Scope of Supply Single Use Circuit- 20 Nos. Trolley- 12 Stage Scavenging Filter – 1 Calibration – 1 set	<b>5.5. To be deleted</b>
Page 8	For Inclusion: 7. Warranty	<b>7. Warranty</b> <b>7.1. Comprehensive Warranty including parts and services for 3 years</b> <b>7.2. Quarterly Preventive Maintenance Services for 6 years</b> <b>7.3. Monthly Calibration Services for 3 years</b> <b>7.4. Calibration Kit goods for 2 years</b>

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 10<sup>th</sup> day of March 2023, BAC Office, PHC.

**MARIETTA A. VELASCO, RN, MAN, MBAH**  
Chairman, BAC for Pharmaceutical Supplies & Med Equipment