



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 023-2022

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 1 unit High End-Echo Machine with 3D/4D and Strain Imaging Capabilities - ITB No. 023.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specification as mentioned above, revisions/changes were hereby introduced.

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
1.2.1 C	Hard drive: at least 4 terabytes (TB)	Hard drive : at least 2 terabytes (TB)
1.6.2	Must submit a certificate that there must be a five-year warranty for parts, labor, transducers, accessories, software and hardware.	Must submit a certificate that there must be a three-year warranty for parts, labor, transducers, accessories, software and hardware.
1.6.5	Must submit a certificate that the vendor shall conduct quarterly preventive maintenance for at least 5 years.	Must submit a certificate that the vendor shall conduct quarterly preventive maintenance for at least 3 years.
1.7.1	Must submit a certificate that the vendor will provide two (2) network switches with 10 Gig SFP+ modules and cable with 48 ports and all Local Area Network (LAN) cables with RJ 45 connectors needed for installation	Must submit a certificate that the vendor will provide two (2) managed network switches with 10 Gig SFP+ modules and cable with 24 ports each and all Local Area Network (LAN) cables with RJ 45 connectors needed for installation (UTP CAT6)
3.5.1	Documentation: Numbering 5.1	Numbering 3.1
3.5.2	Documentation: Numbering 5.2	Numbering 3.2
3.5.2	Must provide a laminated user's guide in English attached to the machine	Must submit a certificate that the vendor will provide a laminated user's guide in English attached to the machine
5.3	For this purpose, contracts similar to the project shall be: a. High-End echo machine with 3D/4D and strain Imaging Capabilities. b. completed within the past 2 years prior to the deadline for the submission and receipt of bids.	For this purpose, contracts similar to the project shall be: a. High-End echo machine with 3D/4D and strain Imaging Capabilities or any ultrasound machine project b. completed within the past 3 years prior to the deadline for the submission and receipt of bids.
	Exclusive Distributor	Exclusive or Authorized Distributor
	Delivery Period- 60 calendar days	90 days calendar days

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 28th day of April 2022, PHC.

ANTONIO D. PASCUAL, MD.

Vice-chairman, BAC for Pharmaceutical Supplies & Med Equipment