



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 025-2023

SUBJECT : Revisions in the Technical Specifications

PROJECT/S : 1 unit Musculoskeletal Ultrasound - ITB No. 030.23

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents, thus, revisions/changes are introduced.

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended "Should be read as"
Page 37	Technical Specifications: 1.3. Ultrasound Transducer 1.3.1. Linear 4-18 Mhz, 1.3.2. Curvilinear 1-5 Mhz 1.3.3. Hockeystick 4-15 Mhz	Technical Specifications: 1.3. Ultrasound Transducer 1.3.1. Linear Lower freq: 1-4 Mhz, Upper freq: 15-18 Mhz 1.3.2. Curvilinear 1-5 Mhz 1.3.3. Hockeystick Lower freq: 2-4 Mhz, Upper freq: 15-16 Mhz
Page 38	1.7. Image format/modes 17.5. Needle enhancement	1.7. Image format/modes 17.5. Needle enhancement: visualization of needle upon insertion to body tissue
	5. Standards, Safety and Training 5.4. Comprehensive warranty for 2 years on equipment 5.5. 3 years warranty for service 5.6. Semi-annual preventive maintenance	5. Standards, Safety and Training 5.4. Comprehensive warranty for 2 years on machine 5.5. 3 years warranty for service and parts 5.6. Semi-annual preventive maintenance for 3 years

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 2nd day of May 2023, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharmaceutical Supplies & Med Equipment