



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 028-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 1 unit Cardiac Compressor (Automatic Cardiac Compressor)
- ITB No. 033.21


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 38	1.5 Weight (excluding battery) 1.5.1 9-10 kgs	1.5 Weight 1.5.1 8-10 kgs
Page 38	1.7 Battery 1.7.2. Capacity: 2300 to 2500 mAh	1.7 Battery 1.7.2. Capacity: 2300 to 3000 mAh

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 28th day of June 2021, BAC Office.


MARIETTA A. VELASCO, RN, MAN.
Chairman, BAC for Pharma Supplies and Med Equipment