



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 030-2024

SUBJECT : Revision in the Bidding Documents

PROJECT : 1 unit Single Plane Angiographic Machine- ITB No. 030.24

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended “Should be read as“
Page 10	Delivery Period: 120 calendar days	Delivery Period: 180 calendar days
Page 40	3.1.2.d Head-toe tilt function	3.1.2.d Head-toe tilt optional function
Page 41	3.1.9.a Minimum of 58 inch medical grade monitor inside the examination room	3.1.9.a Minimum of 58 inch medical grade monitor inside the examination room or 4 units of 27 inch medical grade monitor inside the examination room
Page 43	6.8.1 3 years warranty for angiography system and accessory and 3 years warranty for CATH LAB workstation	6.8.1 The warranty period will start once the machine is declared operational and would last for <ul style="list-style-type: none">● 2 years warranty for angiography system● 1 year on 3rd party (Power Injector, Hemodynamic & CATH LAB workstation)

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 12th day of April 2024, BAC Office, PHC.

ANTONIO D. PASCUAL, MD.
Chairman, BAC for Pharmaceutical Supplies and Medical Equipment