



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 031-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 1 unit Treadmill Stress Test Machine - ITB No. 038.21

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications. (Should be read as) TO
Page 37	Technical Specifications- 1. A2. ECG module must have a battery life indicator power status	 1. A2. ECG module must have a battery life indicator power status (bar type or light indicator type)
	C.1. The user weight capacity of the treadmill must BEAT least 500 lbs. or 227 kilograms.	C.1. The user weight capacity of the treadmill must be at least 496 lbs. or 225 kilograms

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 8th day of July 2021, BAC Office.

MARIETTA A. VELASCO, RN, MAN.

Chairman, BAC for Pharma Supplies and Med Equipment