



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 036-2024

SUBJECT : Revised Technical Specifications of the Project

PROJECT : 9 units Stretcher - ITB No. 035.24


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications is hereby introduced.

- **Attached - Revised Technical Specifications - Annex "A"**

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 16 day of May 2024, BAC Office, PHC.


ANTONIO D. PASCUAL, MD.
Chairman, BAC for Pharmaceutical Supplies and Medical Equipment


S P E C I F I C A T I O N S


9 UNITS STRETCHER

Description of Function	
Stretcher is used to transport patient from one place to another and received medical treatment such as IV medication, insertion prior procedure	
Operational Requirements	
System complete with all accessories as specified in the Technical Specifications below	
Technical Specifications	
1	Hydraulic height adjustment
2	Four (4) wheel casters with double locking system 2.1 Wheel diameter: 7"- 8" (non marking rubber is used for wheels to prevent casters from leaving marks on floor)
3	Dual foot pedal control for Trendelenburg and reverse 0 to 20 degrees
4	Stretcher body: made of stainless steel or powder coated materials
5	Patient weight capacity: at least 250 kg.
6	Height adjustment:
	6.1 High position : 60 - 100 cm from the ground (bed frame only)
	6.2 Lowest position: 24-25 cm from the ground (bed frame only)
7	Width : Minimum - 24 inches, Maximum - 36 inches
8	Length: Minimum- 76 inches , Maximum- 80 inches including the plastic bumpers
9	Side rails : fixable / collapsible with double locking system for patient;s safety
10	Power assist for fowler back rest
11	Mattress:
	11.1 washable and waterproof cover to keep mattress from fluid penetration
	11.2 thickness: 3-4 inches
12	With rubber or polymer plastic bumpers
13	Must have the following accessories:
	13.1 Safety belt: Push button buckle belt type (body and lower extremities) Strap type: Coated polyester vinyl, fluid resistant and can be cleaned and sanitized Strap Width: 2- 3 inches Color: Black
	13.2' Detachable tray with strap to hold monitoring devices weighing at least 5 kg.

	13.3 Compartment for patient's belongings
	13.4 Control brake
	13.5 Integrated portable oxygen holder oblique (stainless), 11-15 cm in diameter
	13.6 Adjustable IV pole *height with 2-4 hook, removable or collapsible
	13.7 Urine bag holder
	13.8 Transport handle at head and lower part

	Standards, Safety and Training
14	Manufacturer/Supplier should have ISO 13485 certificate to Quality Standards
15	With 3-year warranty for parts and services
	Documentation
16	User/Technical/Maintenance manuals to be supplied in English in hard and soft copies upon delivery
17	Exclusive distributorship of similar project
18	Delivery: 90 days


ENGR. JOSE B. BARSAGA
 Chief, EMD


JOYETTE ANN F. MINDORO, MPA, RN
 Chief, ER and Ambulatory Care/ MS1