



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 037-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 16 units Electric Bed-Complete Set - ITB No. 031,21


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 38	Technical Specifications-	Technical Specifications
	8. Color: Beige	8. Color: White or Beige
	Overbed Table: B1. Height adjustable: high position: maximum- 44" (111 cm) Low Position: 29" (74 cm)	Overbed Table: B1. Base Table Height of 29 inches (74 cm) adjustable up to 44 inches (111 cm)

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 9th day of July 2021, BAC Office.


MARIETTA A. VELASCO, RN, MAN.
Chairman, BAC for Pharma Supplies and Med Equipment