



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 038-2024**

**SUBJECT : Revision in the Bidding Documents**

**PROJECT : 14 units Electric Bed (Complete Set) - ITB No. 037.24**


\*\*\*\*\*  
To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
Page 39	I. Technical Specifications: 1.1. Hospital Bed Bed Height Adjustment: a. High Position: maximum 90 cm b. Lowest Position: maximum 43 cm	1.1. Hospital Bed Bed Height Adjustment: a. High Position: <b>less than or equal to 90 cm</b> b. Lowest Position: <b>less than or equal to 46 cm</b>
Page 40	1.2.1. Height Adjustable: High Position maximum- 44" (111 cm) Low Position: 29" (74 cm)	1.2.1. Height Adjustable: High Position: <b>less than or equal to 111 cm</b> Low Position: <b>less than or equal to 85 cm</b>
	1.3. Bedside Table: 1.3.4. Wooden body, high pressure laminate for table top with rounded edges	1.3. Bedside Table: 1.3.4. Wooden body, high pressure laminate <b>OR ABS plastic for table top with rounded edges</b>
Page 41	4.6 Warranty – 3 years on parts and service with quarterly preventive maintenance for 3 years	4.6 Warranty – <b>2 years warranty on parts and services with quarterly preventive maintenance free-of-charge for 3 years</b>

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 17<sup>th</sup> day of May 2024, BAC Office, PHC.

  
**ANTONIO D. PASCUAL, MD.**  
Chairman, BAC for Pharmaceutical Supplies and Medical Equipment