

SUPPLEMENTAL BID BULLETIN No. 039-2024

SUBJECT: Revision in the Bidding Documents

PROJECT : 1 unit Hypo-Hyperthermi a Machine - ITB No. 033.24

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are

hereby introduced

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
	I. Technical Specifications:	1. Technical Specifications:
Page 39	 1.1. Hyper/Hyperthermia Machine 1.1.3. Control Panel (touch screen), with LED digital display of the temperature (both patient and water) in Centigrade and Fahrenheit b. Patient Temperature Range: 25°C to 40°C 	 1.1. Hyper/Hyperthermia Machine 1.1.3. Control Panel (touch screen), with LED digital display of the temperature (both patient and water) in Centigrade and Fahrenheit b. Patient Temperature Range: 25°C to 40°C= NO change as the requested temperature is within the requested range.
	1.2.2. Provision of "repair kit" (for accidental prick)	1.2.2. Provision of "repair kit" (for accidental prick) OR replacement of thermal pad within the warranty period with no additional cost

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 17th day of May 2024, BAC Office, PHC.

ANTONIO DE PASCUAL, MD.

Chairman, BAC for Pharmaceutical Supplies and Medical Equipment