



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 042-2022

SUBJECT : Revision in the Bidding Documents

PROJECT : 9 units Stretcher - ITB.045.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specification as mentioned above, revisions/changes are hereby introduced.

ITB No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 10	Delivery Period: 30 calendar days	Delivery Period: 60 calendar days
Page 37	3. Dual foot pedal control for Trendelenburg and reverse 10 to 30 degrees	3. Dual foot pedal control for Trendelenburg and reverse 0 to 20 degrees
	5. Patient weight capacity: 115-250 kg	5. Patient weight capacity: at least 250 kg
	6. Height adjustment: from 55 cm – 100 cm	6. Height adjustment: a. High position : maximum 100 cm b. Lowest position: maximum 24 cm
	11. Mattress: leatherette and washable thickness: 3"- 4"	11. Mattress: a. washable and waterproof cover to keep mattress from fluid penetration b. thickness: 3-4 inches
Page 38	13.1 Safety belt: Metal push button buckle belt type (body and lower extremities)	13.1 Safety belt: Push button buckle belt type (body and lower extremities)
	13.5 Integrated portable oxygen holder vertical (stainless)	13.5 Integrated portable oxygen holder oblique (stainless)
	13.8 Transport handle head and lower part (rubberized)	13.8 Transport handle head and lower part.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 1st day of August 2022, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN, MBA
Chairman, BAC for Pharma Supplies and Med Equipment