



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 042-2024

SUBJECT : Revised Technical Specifications of the Project

PROJECT : 2 units Automated Sealing Machine - ITB No. 045.24

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced.

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
Page 39	I. Technical Specifications: 3.1. Features 3.1.1. Temperature 3.1.1.2. Sealing temperature maximum of 250»°C	1. Technical Specifications: 3.1. Features 3.1.1. Temperature 3.1.1.2. Sealing temperature maximum of 250»°C – to be deleted
	3.3.1. Packaging Materials • Polypropelene pouches • Tyvek pouches	3.3.1. Automatic in feeding of packaging materials
	3.3.2. Compatible with a variety of packaging materials	3.3.2. Compatible with a variety of packaging materials • Polypropelene pouches • Tyvek pouches
	3.3.5. Minimum production of 4800 pouches an hour	3.3.5. Minimum production of 4000 pouches an hour
	3.4.1. Label printers- one (1) device per sealing machine	3.4.1. Label printers- to be deleted
	3.4.3. Biological indicators with incubator- 50 pcs. Biological indicator for steam with one (1) incubator	3.4.3. Biological indicators with incubator- 50 pcs. Biological indicator for steam with one (1) incubator – to be deleted

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 22nd day of May 2024, BAC Office, PHC.

ANTONIO D. PASCUAL, MD.

Chairman, BAC for Pharmaceutical Supplies and Medical Equipment