



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 043-2024

SUBJECT : Revised Technical Specifications of the Project

PROJECT : 81 units Syringe Pump (Smart Pumps) - ITB No. 036.24

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced.

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
Page 40	I. Technical Specifications: 1.14. The Docking System has: 1.4.a. Provision of at least 4 syringe pumps in 1 docking system with charger according to manufacturer's design.	1. Technical Specifications: 1.14. The Docking System has: a. The docking system is included in the project being bid and not for separate procurement. b. The number of the docking systems to be provided will be dependent on the number of syringe pumps it can accommodate. The minimum number of syringe pumps should be 4 in 1 docking systems. However, if the offered docking systems can accommodate more than 4 syringe pumps, the number of docking systems may be reduced. These docking systems along with the syringe pumps are to be distributed to different ICU's.

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 22nd day of May 2024, BAC Office, PHC.

ANTONIO D. PASCUAL, MD.
Chairman, BAC for Pharmaceutical Supplies and Medical Equipment