



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 044-2022

SUBJECT : Revision in the Bidding Documents

PROJECT : 40 units Electric Bed (Complete Set) - ITB.048.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specification as mentioned above, revisions/changes are hereby introduced.

ITB No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 10	Delivery Period: 30 calendar days	Delivery Period: 90 calendar days
Page 37	1.1. Hospital Bed 1.1.1. Size: 36' x 80" sleeping surface	1.1. Hospital Bed 1.1.1. Size: (W) 34' - 36" x (L) 76"- 80" sleeping surface
Page 38	1.1.5. (g). Digital showing Head of Bed indicator (angle elevation gauge)	1.1.5. (g). Digital indicator of angle elevation of Head of Bed
		Additional Specification for Item 1.1. Hospital Bed Bed Height Adjustment: a). High position: maximum 90 cm b). Lowest position: maximum 40 cm

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 1st day of August 2022, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Med Equipment