



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 044-2024

SUBJECT : Revision in the Technical Specifications of the Project

PROJECT : **2 units Heart Lung Machine with Integrated Centrifugal Pump - ITB No. 043.24**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced.

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
Page 24	5.3. For this purpose, contracts similar to the Project shall be: a. Similar to the Project: Heart Lung Machine with Integrated Centrifugal Pump and Heater Cooler Machine	For this purpose, contracts similar to the Project shall be: a. Heart Lung Machine with Integrated Centrifugal Pump and Heater Cooler Machine OR any Hospital Equipment
Page 39	I. Technical Specifications: 3.2.6. Colored coded alarms	I. Technical Specifications: 3.2.6. Colored visual alarms
40	3.5.3. Pressure Sensor (monitoring of six (6) pressure values	3.5.3. Pressure Sensor (monitoring of at least five (5) pressure values
	3.5.5. Temperature Sensor up to four (4) display simultaneously	3.5.5. Temperature Sensor of at least up to three (3) display simultaneously
	3.7. Inclusion of Perfusion Patient Data Monitor	
	3.7.1. At least twenty one (21) inches Full HD Resolution	3.7.1. Between fifteen (15) to twenty five (25) inches Full HD Resolution
	3.18. With four (4) heavy duty rolling wheels with lock system	3.18. With at least four (4) heavy duty rolling wheels with lock system

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 24th day of May 2024, BAC Office, PHC.

ANTONIO D. PASCUAL, MD.

Chairman, BAC for Pharmaceutical Supplies and Medical Equipment