



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 046-2022**

**SUBJECT : Revisions in the Bidding Documents**

**PROJECT : 4 units Infant/Pediatric/Adult Mechanical Ventilators with Adaptive Support Ventilation Mode (ASD) - ITB.053.22**

\*\*\*\*\*

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specification as mentioned above, revisions/changes are hereby introduced.

ITB No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 22	5.3. For this purpose, contracts similar to the Project shall be: a. Infant/Pediatric/Adult Mechanical Ventilators with Adaptive Support Ventilation Mode (ASD)	5.3. For this purpose, contracts similar to the Project shall be: <b>a. Infant/Pediatric/Adult Mechanical Ventilators with Adaptive Support Ventilation Mode (ASD) or any Mechanical Ventilator.</b>
Page 38	1.3. Required Function: 1.3.1. Standard Monitoring j. With built-in end tidal CO2 and SpO2	1.3. Required Function: 1.3.1. Standard Monitoring <b>j. With built-in end tidal CO2 &amp; SpO2 or with stand-alone end tidal CO2 &amp; in one monitoring device.</b>
Page 39	5. Standard, Safety and Training 5.3. Comprehensive technical training for PHC-BIOMED (factory training) and user's training to staff for familiarity with the system.	5. Standard, Safety and Training 5.3. Comprehensive technical training for PHC-BIOMED and user's training to staff for familiarity with the system by a certified training provider.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 11<sup>th</sup> day of August 2022, BAC Office, PHC.

**MARIETTA A. VELASCO, RN, MAN, MBAH**  
Chairman, BAC for Pharma Supplies and Med Equipment