



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 046-2024**

**SUBJECT** : Revision in the Technical Specifications of the Project

**PROJECT** : 15 units Telemetry Monitoring Devices ( 9 units Telemetry TM 80 with Batteries and 6 units Telemetry BP10 with Case) - ITB No. 040.24


\*\*\*\*\*  
To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced.

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
Pages 1 & 39	15 units Telemetry Monitoring Devices ( 9 units Telemetry TM 80 with Batteries and 6 units Telemetry BP10 with Case)	<b>1 Lot Telemetry Monitoring Devices ( 9 units Telemetry TM 80 with Batteries and 6 units Telemetry BP10 with Case)</b>
Page 24	5.3. For this purpose, contracts similar to the Project shall be: a. Similar to the Project: Telemetry Monitoring Device	For this purpose, contracts similar to the Project shall be: a. <b>Telemetry or Cardiac Monitor</b>
Page 40	1.3 NIBP 1.3.2. Range 1.3.2.2. Diastolic: High= 210 mmHG ± mmHg	1.3 NIBP 1.3.2. Range 1.3.2.2. <b>Diastolic: High= 210 mmHg or higher</b>

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 27<sup>th</sup> day of May 2024, BAC Office, PHC.

  
**ANTONIO D. PASCUAL, MD.**  
Chairman, BAC for Pharmaceutical Supplies and Medical Equipment