



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 048-2022

SUBJECT : Revision in the Bidding Documents

**PROJECT : 1 Lot Supply of Labor and Materials for the
Construction of Oxygen Generating Plant - ITB.056.22**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specification as mentioned above, revisions/changes are hereby introduced.

ITB No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 9	Opening of Bids- August 24, 2022, 11:00 AM., DAPA Hall, PHC.	Opening of Bids- August 24, 2022, 10:00 AM., DAPA Hall, PHC.
Page 10	Delivery Period: 180 calendar days	Delivery Period: 240 calendar days
Page 37	3.1.3. Concentration within the range of 99.5% at Double Stage PSA mode and 95% oxygen purity	3.1.3. Concentration within the range of 99.5% at Double Stage PSA mode and 95% -96% at single stage PSA mode.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 16th day of August 2022, BAC Office, PHC.


RANULFO B. JAVELOSA, JR., MD.
Chairman, BAC for Infra, Med Supplies and Services