



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 052-2024

SUBJECT : Modification in the Technical Specifications of the Project

PROJECT : 1 unit Refrigerated Centrifuge- ITB No. 049.24

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
Item No. 3	3.4 1 unit of 12 blood bags x 450 ml capacity 3.5 LED/LCD digital display of 3.5.1 program selection 3.5.2 speed time 3.5.3 RCF in operation 3.6 Brushless motor 3.10 Electronic door interlock for safe operation	3.4 1 unit of 12 blood bags x 450-500 ml capacity 3.5 LED/LCD, digital/touch screen display of 3.5.1 program selection 3.5.2 speed time 3.5.3 RCF in operation 3.6 Brushless motor or it's equivalent 3.10 Manual or Semi-automatic or Automatic door interlock for safe operation
Item No. 6	6.2. Should be FDA approved products	6..2. Should be FDA approved or equivalent to any government approving body.

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 11th day of June 2024, BAC Office, PHC.

for
ANTONIO D. PASCUAL, MD.

Chairman, BAC for Pharmaceutical Supplies and Medical Equipment