



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 053-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : **162 units Multi Purpose Suction Device - ITB No. 068.21**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
3	Technical Specifications: 3.4. Vacuum Pressure 0- 360 mmHG	3.4. Vacuum Pressure 0- 760 mmHG
	3.8.2. Gauge 0-360 mmHG for Adult	3.8.2. Gauge 0-760 mmHG for Adult
6	Standard, Safety and Training 6.1. Manufactures/Supplier should have ISO 13485 certificate to Quality Standard	6.1. Manufactures/Supplier should have ISO 13485 certificate to Quality Standard (for suction device)
	6.2. Should be FDA (Food and Drug Administration) CF (CF marking is an administrative marking with which the manufacturer or importer affirms its conformity with European health, safety, and environmental protection standards for products sold within the European Economic Area. It is not a quality indicator or a certification mark), UL (Underwriter Laboratories) or BIS (BIS Bureau of Indian Standards) approved product.	6.2. Should be FDA (Food and Drug Administration) CE (CE marking is an administrative marking with which the manufacturer or importer affirms its conformity with European health, safety, and environmental protection standards for products sold within the European Economic Area. It is not a quality indicator or a certification mark), UL (Underwriter Laboratories) or BIS (BIS Bureau of Indian Standards) approved product.


This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 2nd day of November 2021, BAC Office.

RANULFO B. JAVELOSA, JR., M.D.

Chairman, BAC for Infrastructure, Med Supplies and Services


Joyette Ann F. Mindoro
Lic. No. : 0346875