



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 055-2024

SUBJECT : Modification in the Technical Specifications of the Project

PROJECT : 1 unit Tissue Processor - ITB No. 055.24


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced.

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
10	Delivery within 60 calendar days	As is
Page 39	Technical Specifications: 3.4. With reagent cover and charcoal enhanced ventilation to help control processing vapors for user safety.	3.4. With reagent cover and charcoal enhanced ventilation or its equivalent to help control processing vapors for user safety.
	5. Environmental Factors 5.3. Shall meet IEC-60601-1-2-2001 (or equivalent BIS) General requirements of Safety for Electromagnetic Compatibility or should comply with 89/366/EEG; EMC-directive	To be Deleted
	6. Standard, Safety and Training 6.2. Must be local FDA Approved	6.2. Must be local FDA Approved if applicable (CMDR/CMDN/CMDL) and LTO

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 25th day of June 2024, BAC Office, PHC.


ANTONIO D. PASCUAL, MD.
Chairman, BAC for Pharmaceutical Supplies and Medical Equipment