



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 056-2024**

**SUBJECT : Modification in the Technical Specifications of the Project**

**PROJECT : 60 units BP Apparatus with Cuff (Non-Mercurial) - ITB No. 055.24**

\*\*\*\*\*


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced.

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
Page 24	5.3. For this purpose, contracts similar to the Project shall be: a. Similar to the Project: BP Apparatus	5.3. For this purpose, contracts similar to the Project shall be: <b>a. Similar to Project or any medical hospital equipment</b>
Page 39	3. Technical Specifications: 3.5. Type: Wall Mounted	3. Technical Specifications: <b>3.5. Type: Wall Type- 20 units Stand Type- 40 units</b>
	3.6. Square-shaped 14 cm diameter dial	3.6. Square-shaped <b>14 cm-17 cm diameter dial</b>
	3.11. Cuff Mercurial: Deluxe Nylon ( 3 pcs/per unit)	<b>To be deleted</b>
		<b>Additional Requirements:</b> <b>1. Certification of availability of parts &amp; accessories in the market within 5-10 years</b> <b>2. Certification of calibration- Passed</b> <b>3. Preventive Maintenance Schedule (semi-annual)</b> <b>4. Warranty on parts &amp; services</b> <b>5. FDA Approved</b>

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 27<sup>th</sup> day of June 2024, BAC Office, PHC.

  
**ANTONIO D. PASCUAL, MD.**  
Chairman, BAC for Pharmaceutical Supplies and Medical Equipment