



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 057-2024**

**SUBJECT :** Modification in the Technical Specifications of the Project

**PROJECT :** 4 units Cardiac Monitor with Defibrillator - ITB No. 056.24


\*\*\*\*\*  
To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced.

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
Page 40	3. Technical Specifications: 3.4.2.4. Perfusion Indicator Range- 0.05-20%	3. Technical Specifications: 3.4.2.4. <b>To be deleted</b>
	3.4.3.2. Range Systolic: Low- 25 mmHg $\pm$ 5 mmHg High- 290 mmHg $\pm$ 5 mmHg  Diastolic: Low- 10 mmHg $\pm$ 5 mmHg High- 210 mmHg $\pm$ 5 mmHg	<b>Measurement Range: 0 mmHg to 250 mmHg</b>
	3.12 Defibrillation of at least 200 shocks minimum of 200 joules at fully charged battery	<b>3.12 Defibrillation of at least 200 shocks minimum of 200 joules at fully charged battery or its equivalent</b>

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 27<sup>th</sup> day of June 2024, BAC Office, PHC.

  
**ANTONIO D. PASCUAL, MD.**  
Chairman, BAC for Pharmaceutical Supplies and Medical Equipment