



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 061-2022

SUBJECT : Revisions in the Technical Specification

PROJECT/S : 1 unit Musculoskeletal Ultrasound - ITB.071.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
37	1. Technical Specification: 1.2. At least 24” full HD LED monitor with high quality display and ensure best image quality and highest frame rate	1. Technical Specification: 1.2. At least 21” full HD LED monitor with high quality display and ensure best image quality and highest frame rate
	1.4. Ultrasound transducer 1.4.1. Linear (6-19 Mhz) 1.4.2. Curvilinear (2.5 Mhz) 1.4.3. Hockeystick (6-18 Mhz)	1.4. Ultrasound transducer 1.4.1. Linear (8-15 Mhz) 1.4.2. Curvilinear (1-6 Mhz) 1.4.3. Hockeystick (4-15 Mhz)
38	2. System Configuration Accessories, spares and consumables	2. System Configuration Accessories, spares and consumables 2.3. Initial Stocks 2.3.1. (2) Gal ultrasound gel 2.3.2. (5) rolls thermal paper
	6. Documentation 6.4. Delivery: 30 days	6. Documentation 6.4. Delivery: 60 days

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 30th day of September 2022, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Medical Equipment