



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 062-2022

SUBJECT : Revision in the Technical Specification

PROJECT/S : 1 unit Treadmill Stress Testing Machine - ITB.069.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revision/change is hereby introduced.

Pasge No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
38	1. 3. Treadmill: 1.3.1. The user weight capacity of the treadmill must be at least 500 lbs or 227 kilograms.	1. 3. Treadmill: 1.3.1. The treadmill must accommodate a maximum user weight of 500 lbs or 227 kilograms.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 30th day of September 2022, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Medical Equipment

