



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 070-2022

SUBJECT : Modifications/Amendments in the Bidding Documents

**PROJECT/S : Multi-Year Project for the Urine Analyzer
(Clinical Microscopy Section) – ITB No. 084.22**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
B	Specification for Urine Analyzer:	
6	With at least 10 parameters for urine chemistry strips	<i>More than</i> 10 parameters for urine chemistry strips
7	Capable of body fluid analysis specific for pH and specific gravity	Capable of body fluid analysis specific for pH and specific gravity (<i>Optional</i>)
8	With at least 2-3 ml sample volume in tube	With 2-3 ml sample volume <i>or less</i> in tube
C	GENERAL REQUIREMENTS	
9	Provide more or less 1000 pieces urine test tubes per month to be used during analysis	Provide more or less 1000 pieces <u>10 ml</u> urine test tubes per month to be used during analysis
	Additional Requirement:	MUST Provide Certification of Assurance of Continuous Supply of Reagents, Controls and Consumables from the Principal and Distributor for Three Years

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 19th day of October 2022, BAC Office, PHC.


RANULFO B. JAVELOSA, JR., MD.
Chairman, BAC for Infra, Med Supplies and Services