



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 071-2022

SUBJECT : Modifications/Amendments in the Bidding Documents

PROJECT/S : Medical Supplies (Early Procurement Activity for CY 2023)

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
A). - Attached Revised Selected Medical Supplies as previously pre-issued in the Bidding Documents		
B). Page 39	D. DELIVERIES AND PENALTIES 3. All deliveries shall be subject to the following conditions: 3.5. Supplier shall assure the PHC of a continuous supply of all items awarded. In the event cases the Suppliers fail to deliver the items awarded, PHC shall be free to buy its requirements from other sources and price different based on Contract price will be chargeable against the major awardee.	D. DELIVERIES AND PENALTIES 3. All deliveries shall be subject to the following conditions: 3.5. Principal and Distributor MUST provide Certificate of Assurance to PHC of a continuous supply of all items awarded for One (1) year. In the event that the distributor fails to deliver the items awarded, PHC shall be free to buy its requirements from other sources and price different based on Contract price will be chargeable against the major awardee.
	Changes in the Approved Budget for the Contract (ABC): 1. OR Supplies = P179,958,447.91 2. Medical/Surgical Supplies = P105,639,379.48 3. CSS Supplies = P267,876,208.36 4. Miscellaneous Supplies = P199,341,204.62 ABC = P742,714,240.37	Changes in the Approved Budget for the Contract (ABC): 1. OR Supplies = P179,048,284.81 2. Medical/Surgical Supplies = P105,597,412.20 3. CSS Supplies = P259,122,931.76 4. Miscellaneous Supplies = P198,944,006.42 ABC = P742,712,635.19

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 19th day of October 2022, BAC Office, PHC.


RANULFO B. JAVELOSA, JR., MD.
Chairman, BAC for Infra, Med Supplies and Services

**PHILIPPINE HEART CENTER
East Avenue, Quezon City**

Revision in the Abstract of Sealed Bidding: Medical Supplies, for the period of January 2023 to December 2023 (ITB.085.22)

BID FORM

Item#	SPECIFICATION	UNIT OF MEASURE	TOTAL QTY	UNIT COST	BID PRICE	BIDDER
OR SUPPLIES						
Note : PERCUTANEOUS ENDOSCOPIC GASTROSCOPY (PEG) SET (DELETED AND REPLACED WITH ITEM BELOW)						
25	INCISE DRAPE, LONG, CLEAR, WITHOUT POVIDONE, 56 CM X 84 CM	PIECE	688	1,774.50		
64	Isolation Gown (OR and CVLab Use)	PACK	18,000	148.58		
	1. Medical grade					
	2. Non-woven material					
	3. Non-sterile					
	4. With AAMI (Association for the Advancement of Medical Instrumentation) PB70:2012 certification(at least level 3 barrier protection or its equivalent)					
	5. Color- Blue					
	6. Long sleeved with elastic wrist					
	7. Ideally with thumb hole design					
	8. Individually packed					
	9. Dimensions:					
	9.1 Width – 70 cm (+/- 5 cm)					
	9.2 Length -115 cm (+/- 5 cm)					
	9.3 Arm hole – 25 cm (+/- 5 cm)					
MEDICAL/SURGICAL SUPPLIES						
67	DISPOSABLE TRANSDUCER KIT (MUST BE COMPATIBLE WITH EXISTING MACHINE)	PIECE	9,900	1,401.75		
73	ACT CUVETTES NOTE: MUST BE COMPATIBLE WITH EXISTING MACHINE	box	400	14,508.18		
81	IV CATHETER, G-24 with safety features can be used for ADULTS	PIECE	59,400	121.80		
82	PREMIUM PEDIA IV CATHETER G-24 selective use of CSS	PIECE	52,800	83.77		
CSS SUPPLIES						
99	DISPOSABLE EXAMINATION GLOVES,100'S, MEDIUM, NON-STERILE, POWDER FREE	BOX of 100	39,600	325.50		
144	ENZYMATIC CLEANER (MULTI), 4 TO 5 LITERS	GAL	660	6,812.40		
154	INCISE DRAPE, SHORT, CLEAR, WITHOUT POVIDONE 34-55CM x 35-45 CM	PIECE	2,376	861.00		
157	BLOOD GAS AND ELECTROLYTE CARTRIDGES must provide machine;	PIECE	26,246	501.90		
	TECHNICAL SPECIFICATIONS FOR CARTRIDGES					
	1. Test must provide results for:					
	a. Blood Gas (pH, pCO ₂ , pO ₂ , HCO ₃ , TCO ₂ , PE, O ₂ Sat					
	b. Electrolytes (Na, K, iCa)					
	c. Hematology					
	d. Glucose					
	e. Lactate					
	2. The distributor will replace any defective cartridges that is being delivered to the end user					
	TECHNICAL SPECIFICATIONS FOR ABG LOANER UNITS					
	1. The machine must be capable of determining the following tests:					
	a. Blood Gas					

Philippine Heart Center
East Ave., Quezon City
APPROVED

[Signature]
Bid-User's Signature

	b. Electrolytes and Hematology				
	c. Chemistry				
	d. Coagulation				
	e. Cardiac Marker				
	2. Keypad control (use of finger only and not stylus)				
	3. Capable of linking with data management system thru USB or network				
	4. Reading time for test must not exceed 3 minutes				
	5. Must be battery operated				
	6. The distributor will provide a battery replacement to ensure a continuous machine operation				
	7. Must have at least LCD screen				
	8. The weight must not exceed 1.5 lbs				
	9. Must provide at least four (4) ABG machines				
	10. Must include printer for result printing				
	11. Maintenance and replacement of parts should be the sole responsibility of the distributor				
	12. Must perform a free regular preventive maintenance checkup				
	13. The distributor must provide certificate of training of the technician who will perform the maintenance checkup				
167	I.V. EXTENSION TUBING 60 TO 75 CM, VOLUME (40 deg C, 2 BAR) 6.3 ml/meter	PIECE	19,800	325.50	
168	I.V. EXTENSION TUBING 140 TO 150 CM, VOLUME (40 deg C, 2 BAR) 6.3 ml/meter	PIECE	39,600	61.74	
176	TRANSPARENT FILM DRESSING W/ NON-ADHESIVE PAD, STERILE, 6 CM TO 8 CM X 10 CM, STERILE	BOX OF 50	1,056	1,774.50	
183	TRANSPARENT FILM DRESSING WITH NON-ADHESIVE PAD, 8CM TO 9CM X 15CM, STERILE	PIECE	39,600	73.46	
	MISCELLANEOUS SUPPLIES				
199	ANTI A TYPING SERA 10 ML.	VIAL	900	472.50	
200	ANTI B TYPING SERA 10 ML.	VIAL	900	472.50	
201	ANTI HUMAN SERUM , GLOBULIN 10 ML	VIAL	900	772.80	
202	ANTI RH / D 10 ML	VIAL	600	633.15	
203	Antibody Enhanced Solution (LISS) 10 ml Note : 1 lot for item Nos. 199 to 203 All or None Policy, Valid FDA registration required	VIAL	600	653.10	
217	CATHETER SHEATH INTRODUCER SYSTEM, FR.5 FEMORAL 10'S Kit includes: Cannula: Usable length 11 cm Mini guidewire Diameter: 0.038, Length 45 cm	PIECE	1,320	840.00	
218	CATHETER SHEATH INTRODUCER SYSTEM, 6F, FEMORAL, 10'S Kit includes: Cannula: Usable length 11 cm Mini guidewire Diameter: 0.038, Length 45 cm	PIECE	3,960	840.00	
219	CATHETER SHEATH INTRODUCER SYSTEM, FR7, FEMORAL 10's Kit includes: Cannula: Usable length 11 cm Mini guidewire Diameter: 0.038, Length 45 cm	PIECE	3,960	840.00	
220	CATHETER SHEATH INTRODUCER SYSTEM, Fr.08, FEMORAL, 10's Kit includes: Cannula: Usable length 11 cm Mini guidewire Diameter: 0.038, Length 45 cm	PIECE	1,060	840.00	
222	INTRA AORTIC BALLOON CATHETER FR. 8 X 30 capacity (CV Lab Use) Kit includes	PIECE	158	38,955.00	

Philippine Heart Center
East Ave., Quezon City
APPROVED

[Signature]
Head-User's Signature

	1: 8Fr. (2.7mm) – 30 cc IAB Catheter with Pre-Mounted Peel-Away Hemostasis Device				
	1: Driveline Tubing with Pre-Attached IAB console Connector				
	1: 60ml Syringe				
	1: 8Fr. (2.7mm) x 6" (15cm) PTFE Hemostasis Sheath				
	Introducer with Dilator Assembly				
	1: 8Fr. (2.7mm) x 6" (15cm) PTFE Hemostasis Sheath Introducer with Dilator Assembly				
	2: .025" (0.64mm) x 69" (175cm) PTFE Coated "J" Guidewires				
	1: 18 Ga. (1.4mm) Arterial Needle				
	1: Pre-Dilator				
	1: #11 Scalpel				
	1: 86" (218 cm) Driveline Tubing Adapter				
223	INTRA AORTIC BALLOON CATHETER, FR. 8 , 40cc capacity	PIECE	264	39,039.00	
	Kit includes 1: 8Fr. (2.7mm) – 40 cc IAB Catheter with Pre-Mounted Peel-Away Hemostasis Device 1: Driveline Tubing with Pre-Attached IAB console Connector 1: 60ml Syringe 1: 8Fr. (2.7mm) x 6" (15cm) PTFE Hemostasis Sheath Introducer with Dilator Assembly 1: 8Fr. (2.7mm) x 6" (15cm) PTFE Hemostasis Sheath Introducer with Dilator Assembly 2: .025" (0.64mm) x 69" (175cm) PTFE Coated "J" Guidewires 1: 18 Ga. (1.4mm) Arterial Needle 1: Pre-Dilator 1: #11 Scalpel 1: 86" (218 cm) Driveline Tubing Adapter				

Philippine Heart Centre
East Ave., Quezon City
APPROVED

[Handwritten Signature]
End-User's Signature