



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 074 -2022

SUBJECT : Modifications/Amendments in the Bidding Documents

PROJECT/S ; 1 pc. Dual Energy Multi-Slice CT Scan machine with capabilities for cardiovascular brain, Head and Neck and Body imaging including all peripherals necessary for these examination and for the safety of patients and those who will perform the imaging studies as well as the necessary capabilities for storage and transmittal of imaging data to the existing PACS of the PHC - ITB No. 089.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
10	Approved Budget for the Contract (ABC) was P150,000,000.00 as posted and advertised.	Approved Budget for the Contract (ABC) was reduced to P130,000,000.00 due to the inquiry that the equipment/unit present market value.
	Delivery Period - 60 calendar days	Delivery Period - 120 calendar days
22	Contract Similar to the Project for CT-MRI	Single Largest Contract for CT-MRI (2019 onward)
38	1.3.1. Detector Configuration – Multi Layer	1.3.1. Detector Configuration – Multi Layer or equivalent configuration (Spectral Acquisition)
39	1.6.8. Spectral Applications b. Interactive Reconstruction Technique - Required	1.6.8. Spectral Applications b. Iterative Reconstruction Technique - Required
	1.7.1. Manufacturer's Standard Latest version (High Performance Computer) - Required	1.7.1. Manufacturer's Standard Latest version (High Performance Computer) – Manufacturer's Standard Specifications
	1.7.2. Console Table – Required	1.7.2. Console Table – Manufacturer's Standard Specifications
	1.7.3. Keyboard – Required	1.7.3. Keyboard – Manufacturer's Standard Specifications
	1.7.4. Mouse – Required	1.7.4. Mouse – Manufacturer's Standard Specifications
	1.7.5. Two (2) units LCD Monitor- (4-5 megapixels)- at least 27" (Image Display and Control Desk Display for all parameters)	1.7.5. Two (2) units LCD Monitor- (4-5 megapixels) - Manufacturer's Standard Specifications
	1.7.6. CPU - a). RAM 32GB or Higher b). Hard Disk Capacity 5TB or Higher	1.7.6. CPU - a). Manufacturer's Standard Specifications b). Manufacturer's Standard Specifications

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
39	c). Processor at Least Quadcone	c). Manufacturer's Standard Specifications
	d). Additional Archiving Storage CD/DVD/USB	d). Manufacturer's Standard Specifications
41	1.11.28. Ion Detection Technology - Required	1.11.28. Ion Detection Technology – Required or Equivalent Technology for Tissue Characterization
42	2.18. Infant Cradle without Metal Parts - Required	2.18. Infant Immobilization System without Metal - Required
43	5.5. Warranty/After Support 5.5.1. All Parts and Services - 5 years	Reduced to 3 years warranty
44	Documentation:	Additional Requirement: Duly Notarized Certificate from the manufacturer that they have a Physical Office/service center in the Philippines.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 20th day of October 2022, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Medical Equipment