



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 075 -2022

SUBJECT : Modifications/Amendments in the Bidding Documents

PROJECT/S : 8 units Infant/Pediatric Mechanical Ventilators with Adaptive Support Ventilation Mode (ASV)- ITB No. 088-2022

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 37	Technical Specification: 1.2.1. Adaptive Support Ventilation (ASV), Adaptive Ventilation Mode (AVM) or it's equivalent	1.2.1. Adaptive Support Ventilation (ASV) or Adaptive Minute Ventilation (AMV), or Adaptive Ventilatory Assist (AVA) or Adaptive Pressure-base intelligent Ventilation Support (AdVent)
Page 38	1.2.1. g. Pressure Flow	1.3.1. g. Pressure and Flow
	1.3.1.h. Pressure Flow and Pressure/Volume loops	1.3.1.h. Pressure Flow and Pressure/Volume loops
	1.3.4. Standard Settings e. Pressure control 0-100 cmH2O f. Pressure support 0-100 cmH2O	1.3.4. Standard Settings e. Pressure control lower limit of 0 to 20 cmH2O upper limit 60-100 cmH2O f. Pressure support lower limit of 0 to 2 cmH2O upper limit between 60-100 cmH2O
Page 39	5. Standard, Safety and Training 5.3. Comprehensive technical training for PHC-BIOMED (factory training) and User's training to staff for familiarity with the system	5. Standard, Safety and Training 5.3. Comprehensive technical training for PHC-BIOMED and End-user's for familiarity with the system by a certified training provider

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 21st day of October 2022, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Medical Equipment