



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 076 -2022

SUBJECT : Modifications/Amendments in the Bidding Documents

PROJECT/S ; 1 unit Portable Surface Ultrasound - ITB No. 091.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 10	Delivery Period – 30 calendar days	Delivery Period – 60 calendar days
Page 37	Technical Specification: 2.1. Ultrasound guidance for vascular access procedures should be performed during broadband high-resolution linear array transducers 2.1.1. Frequencies between 7.5 and 12 MHz is preferred 2.1.2. Single probe that will cover this frequency adjustable for high frequency viewing superficial vessels 2.1.3. This probe can also be adjustable to lower frequency for deeper vessel viewing	2.1. Ultrasound guidance for vascular access procedures should be performed during broadband high-resolution linear array transducers 2.1.1. Frequencies should cover from 7.5 and 12 MHz is preferred 2.1.2. Probe that will cover this frequency adjustable for high frequency viewing superficial vessels 2.1.3. Probe can also be adjustable to lower frequency for deeper vessel viewing Additional Requirement: 2.1.4. Portable ultrasound with console
Page 38	Power Supply:	Additional Requirement: 5.2. UPS (Uninterruptible Power Supply) at 2KVA single- phase, 220-240V

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 21st day of October 2022, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Medical Equipment

