



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 080 -2022

SUBJECT : Modifications/Amendments in the Bidding Documents

PROJECT/S : 2 units Direct Radiography Flat Panel Detector 14" x 17" (X-Ray Detector) - ITB No. 106.22


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 38	1.13. Work Station 1.13.1. Desktop (One unit) e. Monitor: 27” Touch Screen Monitor, 32”, 2MP clinical & Medical Grade Monitor	1.13. Work Station 1.13.1. Desktop (One unit) e.1. Monitor: 27” Touch Screen Monitor e.2. 32”, 2MP clinical & Medical Grade
39	2. System Configuration Accessories, spares and consumables 2.1. Five (5) years of free comprehensive maintenance , parts of entire system	2. System Configuration Accessories, spares and consumables 2.1. Three (3) years of free comprehensive maintenance , parts of entire system
	5. Standard, Safety and Training 5.1. Manufactures/Supplier should have ISO 13485 certificates to Quality Standard	5. Standard, Safety and Training 5.1. Manufactures should have ISO 13485 certificates to Quality Standard
	5.5. The warranty period will start once the machine is declared operational and would last for five (5) years	5.5. The warranty period will start once the detector system is declared operational and would last for three (3) years
Page 40	5.8. Maintenance parts needed to be replaced will still fall under warranty period of 3 years	5.8. Maintenance parts needed to be replaced (including battery & cable wire) will still fall under warranty period of 3 years

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 27th day of October 2022, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Medical Equipment

