



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 089-2022**

**SUBJECT : Modifications/Amendments in the Bidding Documents**

**PROJECT/S : 9 units Stretcher (Re-Bid) - ITB No. 115.22**

\*\*\*\*\*

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
6	Height adjustment: a. High position: maximum 100 cm b. Lowest position: maximum 24 cm	Height adjustment: a. High position: maximum 100 cm b. <b>Lowest position: maximum 54 cm</b>

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 2<sup>nd</sup> day of December 2022, BAC Office, PHC.

  
**MARIETTA A. VELASCO, RN, MAN, MBAH**  
Chairman, BAC for Pharma Supplies and Med Equipment