



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 091-2022

SUBJECT : Modifications/Amendments in the Bidding Documents

PROJECT/S : 40 units Electric Bed – Complete Set (Re-Bid) - ITB No. 116.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 37	Height adjustment: a. High position: maximum 90 cm b. Lowest position: maximum 40 cm	Height adjustment: a. High position: maximum 90 cm b. Lowest position: maximum 43 cm
Page 38	1.1.6. b.1. Portable O2 holder: 11.5 cm in diameter	1.1.6. b.1. Portable O2 holder: 11-15 cm in diameter with adjustable screw to secure portable O2 tank
	1.3. Bedside Table 1.3.6. Preferably color-coordinated and same brand with bed (beige)	1.3. Bedside Table 1.3.6. Preferably color-coordinated and same brand with bed (white or beige)

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 2nd day of December 2022, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Med Equipment