



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 092-2022

SUBJECT : Modifications/Amendments in the Bidding Documents

PROJECT/S : 1 unit Hypo/Hyperthermia (Re-Bid) - ITB No. 118.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

| Item No./ Page No. | PHC – Technical Specifications FROM | Clarifications/Modifications (Should be read as) TO |
|-----------------------|---|---|
| | 1.1.3. Control panel (touch screen) with LED digital display of temperature (both patient and water in Centigrade and Fahrenheit b. Patient Temperature Range: 25°C-40°C | 1.1.3. Control panel (pressure button /touch screen) with LED digital display of temperature (both patient and water in Centigrade b. Patient Temperature Range: 30°C-40°C |
| Page 10 | Delivery Period: 30 calendar days | Delivery Period: 90 calendar days |

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 2nd day of December 2022, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Med Equipment