



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 095-2022

SUBJECT : Modifications/Amendments in the Bidding Documents

PROJECT/S : 1 unit Semi-Automated Rotary Microtome - ITB No. 111.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revision/change is hereby introduced.

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 37	Technical Specifications: 3. With waterbath feature (preferred temperature: 40-50 °C	Technical Specifications: 3. With integrated waterbath feature (preferred temperature: 40-50 °C

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 2nd day of December 2022, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Med Equipment