



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 096-2022

SUBJECT : Modifications/Amendments in the Bidding Documents

**PROJECT/S : 2 units Micro Transesophageal Echocardiogram (TEE)
Transducer Probes - ITB No. 119.22**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 37	3. Technical Specifications: 3.1. Capability 2D-4D TEE 3.2. Patient Weight: 2.5 kg – 3.5 kg to 20 kg	3. Technical Specifications: 3.1. Capability 2D TEE 3.2. Patient Weight: 2.5 kg – 20 kg (for 1 unit) and 3.5 kg - 20 kg (for 1 unit)

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 6th day of December 2022, BAC Office, PHC.

RANULFO B. JAVELOSA, JR., MD.
Chairman, BAC for Infra, Med Supplies and Services